

Red Liquid Masons Hygrometer - 736740  
Difference between DRY BULB and WET BULB readings (Celcius)

Dry Bulb Temp.	0.5	10.0	1.5	2.0	2.5	3.0	3.5	4.0	45.0	5.0	5.5	6.0	6.5	7.0	7.5	8.0
50	97	94	92	89	87	84	82	79	77	74	72	70	68	66	63	61
49	97	94	92	89	86	84	81	79	77	74	72	70	67	65	63	61
48	97	94	92	89	86	84	81	79	76	74	71	69	67	65	62	60
47	97	94	92	89	86	83	81	78	76	73	71	69	66	64	62	60
46	97	94	91	89	86	83	81	78	76	73	71	68	66	64	62	59
45	97	94	91	88	86	83	80	78	75	73	70	68	66	63	61	59
44	97	94	91	88	86	83	80	78	75	72	70	68	65	63	61	58
43	97	94	91	88	85	83	80	77	75	72	70	67	65	62	60	58
42	97	94	91	88	85	82	80	77	74	72	69	67	64	62	59	57
41	97	94	91	88	85	82	79	77	74	71	69	66	64	61	59	56
40	97	94	91	88	85	82	79	76	73	71	68	66	63	61	58	56
39	97	94	91	87	84	82	79	76	73	70	68	65	63	60	58	55
38	97	94	90	87	84	81	78	75	73	70	67	65	62	59	57	54
37	97	93	90	87	84	81	78	75	72	69	67	64	61	59	56	54
36	97	93	90	87	84	81	78	75	72	69	66	63	61	58	55	53
35	97	93	90	87	83	80	77	74	71	68	65	63	60	57	55	52
34	96	93	90	86	83	80	77	74	71	68	65	62	59	56	54	51
33	96	93	89	86	83	80	76	73	70	67	64	61	58	56	53	50
32	96	93	89	86	83	79	76	73	70	67	64	61	58	55	52	49
31	96	93	89	86	82	79	75	72	69	66	63	60	57	54	51	48
30	96	93	89	85	82	78	75	72	68	65	62	59	56	53	50	47
29	96	92	89	85	81	78	74	71	68	65	61	58	55	52	49	46
28	96	92	88	85	81	77	74	70	67	64	60	57	54	51	48	45
27	96	92	88	84	81	77	73	70	66	63	60	56	53	50	47	44
26	96	92	88	84	80	76	73	69	66	62	59	55	52	49	46	42
25	96	92	88	84	80	76	72	68	65	61	58	54	51	47	44	41
24	96	91	87	83	79	75	71	68	64	60	57	53	50	46	43	39
23	96	91	87	83	79	75	71	67	63	59	56	52	48	45	41	38
22	95	91	87	82	78	74	70	66	62	58	54	51	47	43	40	36
21	95	91	86	82	78	73	69	65	61	57	53	49	45	42	38	35
20	95	91	86	81	77	73	68	64	60	56	52	48	44	40	36	33
19	95	90	86	81	76	72	67	63	59	55	50	46	42	38	34	31
18	95	90	85	80	76	71	66	62	58	53	49	45	41	36	32	29
17	95	90	85	80	75	70	65	61	56	52	47	43	39	34	30	26
16	95	89	84	79	74	69	64	60	55	50	46	41	37	32	28	24
15	94	89	84	78	73	68	63	58	53	49	44	39	35	30	26	21
14	94	89	83	78	72	67	62	57	52	47	42	37	32	28	23	18
13	94	88	83	77	71	66	61	55	50	45	40	35	30	25	20	16
12	94	88	82	76	70	65	59	54	48	43	38	32	27	22	17	12
11	94	87	81	75	69	63	58	52	46	41	35	30	25	19	14	9
10	93	87	81	74	68	62	56	50	44	38	33	27	22	16	11	5
9	93	86	80	73	67	61	54	48	42	36	30	24	18	13	7	2
8	93	86	79	72	66	59	52	46	40	33	27	21	15	9	3	
7	93	85	78	71	64	57	50	44	37	31	24	18	11	5		
6	92	85	77	70	63	55	48	41	34	28	21	14				
5	92	84	76	69	61	53	46	39	31	24						
4	92	83	75	67	59	51	44	36								
3	91	83	74	66	57	49										
2	91	82	73	64												
1	90	81														

Fill the reservoir and ensure the wick is wet Note the Dry Bulb reading (left hand column). Subtract the Wet Bulb read, from the Dry Bulb reading to find the Wet Bulb depression. Note the difference along the top line. Follow this column down until level with the Dry Bulb reading for the Percentage Humidity. Best results will be achieved when the instrument is situated in slightly moving air (approx 1~15 meters/sec). Dimensions: 12 x 2-7/8 x 1-1/4", Weight: 3.8 oz.